



Performance Evaluations of Instruments Designed for Rapid, Shipboard Detection of Living Microorganisms in Ballast Water

Matthew R. First¹, Vanessa Molina², Stephanie H. Robbins-Wamsley²,
Scott C. Riley², Cameron S. Moser³, Mario N. Tamburri⁴,
Thomas H. Johengen⁵, Heidi Purcell⁵, G. Jason Smith⁶,
Earle N. Buckley⁷, and Lisa A. Drake³

¹Code 6137, Naval Research Laboratory, Washington, DC 20375

²Excet, Inc.; Springfield, VA 22150

³Code 6137, Naval Research Laboratory, Key West, FL 33041

⁴University of Maryland Center for Environmental Science; Solomons, MD 20688

⁵Cooperative Inst. For Limnology and Ecosystems Res.; Ann Arbor, MI 48109

⁶Moss Landing Marine Lab.; Moss Landing, CA 95039


⁷Buckley Environmental, Mount Pleasant, SC 29464



Background on “Compliance Tools”

Compliance Tools:

- Test ballast water as it is discharged
- Designed for rapid, shipboard analysis
- Typically report risks of exceeding the 10 mL^{-1} limit for organisms ≥ 10 and $< 50 \mu\text{m}$



A Framework for Validation*

Step 1: Proof-of-Concept

- Pilot study
- Subject matter workshops

Step 2: Verification and Validation

- Rigorous, independent testing
- Tests with challenging conditions

Step 3: Feasibility and Selection

Considerations include:

- Functional requirements
- Physical size and safety
- Cost and ease-of-use

2015 → 2016:

Testing of compliance
tools based upon
variable fluorescence
fluorometry

*Drake et al. (2014) Marine Pollution Bulletin 86: 122-128




Required Method for Organisms $\geq 10 \mu\text{m}$ and $< 50 \mu\text{m}$

The **Environmental Technology Verification Protocol (ETV)*** stipulates an approach based upon epifluorescence microscopy





Step 1: Labeling

Two fluorescent probes are introduced into the sample



Step 2: Manual microscopy

Visual counts of fluorescing or moving (i.e., living) organisms



*U.S. Environmental Protection Agency, 2010; the ETV is the U.S. protocol for land-based verification testing of ballast water management systems

Laboratory and Field Trials

Laboratory trials:

Tested a range of concentrations of one of two cultured microalgae



Prorocentrum micans *Tetraselmis marina*

Target concentrations:

0 mL ⁻¹	20 mL ⁻¹
5 mL ⁻¹	50 mL ⁻¹
10 mL ⁻¹	100 mL ⁻¹

Field trials:

Examined ambient samples at contrasting locations



Fluorometry-based compliance tools

YSI Ballast Monitor
Xylem
86 x 103 x 30 cm
100 kg



10Cells
BBE Moldaenke
30 x 34 x 15 cm
5 kg



BW680
Hach
6 x 14 x 5 cm
0.3 kg

Ballast-Check 2
Turner Designs
9 x 18 x 5 cm
0.4 kg



FastBallast
Chelsea Tech. Group
20 x 24 x 5 cm
3 kg

Testing and Analysis: 2015-2016

Round 1: June – September, 2015

Round 2: March – July, 2016



Sampling ambient organisms
from seawater in Key West, FL

Linearity

- Do measurements of abundance change proportionately with cell concentrations?

Precision

- Are repeated measurements of the same sample in agreement?

Accuracy

- Does the instrument's assessment (i.e., above or below the discharge standard) agree with microscope counts?

Results: Linearity (All trials)

R^2 Values: Coefficient of Determination
Microscope counts vs. compliance tool concentrations

Tool	Laboratory Trials			Field Trials			
	<i>T. marina</i>	<i>P. micans</i>	Both	NRL	GSI	SERC	All Sites
Ballast-Check 2 (2015)	0.46	0.98	0.90	0.63	0.64	0.12	0.10
10Cells	0.85	0.84	0.68	0.61	0.69	0.68	0.48
YSI Ballast Monitor	0.87	0.94	0.91	0.72	0.66	0.01	0.15
Ballast-Check 2 (2016)	0.33	0.90	0.82	0.73	0.46	0.39	0.36
FastBallast	N/A	N/A	N/A	0.13	0.75	0.71	0.37
BW680	0.57	0.92	0.86	0.66	0.61	0.82	0.66

R^2 Values: 0 to 1

Detailed reports available at: www.act-us.info

Legend:

$R^2 \geq 0.90$ $R^2 \geq 0.75$

$R^2 < 0.50$

Results: Precision (Laboratory trials)

CV: Coefficient of Variation

Tool	Minimum	Maximum	Mean	Median	n
BallastCheck2 (2015)	22%	230%	77%	59%	21
10Cells	N/A	N/A	N/A	N/A	N/A
YSI Ballast Monitor	0.2%	24%	4.7%	3.4%	36
BallastCheck2 (2016)	1%	99%	33%	29%	14
FastBallast	N/A	N/A	N/A	N/A	N/A
BW680	2%	105%	30%	16%	23

CV (%): Standard deviation adjusted to the mean
Only reported for mean values >10 units

Legend:
CV <25%
CV ≥25%

Results: Precision (Field trials)

CV: Coefficient of Variation

Tool	Minimum	Maximum	Mean	Median	n
BallastCheck2 (2015)	9%	61%	28%	26%	12
10Cells	6%	52%	24%	22%	20
YSI Ballast Monitor	0.1%	63%	13%	4.7%	36
BallastCheck2 (2016)	25%	113%	63%	53%	15
FastBallast	9%	42%	21%	18%	22
BW680	6%	101%	25%	17%	26

CV(%): Standard deviation adjusted to the mean
Only reported for mean values >10 units

Legend:
CV <25%
CV ≥25%

Results: Accuracy (Laboratory trials)

Probability of measuring an exceedance at 30 mL⁻¹

Compliance Tool	Laboratory Trials		
	<i>T. marina</i>	<i>P. micans</i>	Both organisms
Ballast-Check 2 (2015)	0.62	0.98	0.71
10Cells	N/A: Insufficient readings exceeding 10 mL ⁻¹		
YSI Ballast Monitor	N/A: Pass/Fail not reported		
Ballast-Check 2 (2016)	N/A ²	0.99	0.64
FastBallast	N/A: Instrument malfunction		
BW680	1.00	1.00	0.99

30 mL⁻¹: 3x the exceedance of the discharge standard

Legend:
Probability ≥ 0.90

Results: Accuracy (Field trials)

Probability of measuring an exceedance at 30 mL⁻¹

Compliance Tool	Field Trials			
	NRL	GSI	SERC	All Sites
Ballast-Check 2 (2015)	0.97	0.26	0.07	0.28
10Cells	0.99	0.98	1.00	0.99
YSI Ballast Monitor	N/A: Pass/Fail not reported			
Ballast-Check 2 (2016)	1.00	0.22	0.05	0.25
FastBallast	N/A: Insignificant regression			0.70
BW680	1.00	1.00	0.99	0.97

30 mL⁻¹: 3x an exceedance of the discharge standard

Legend:
Probability ≥ 0.90

Conclusions: Testing the Validation Framework

Tests provided challenging conditions, and in general, the compliance tools performed well for samples of:

- Monocultures of relatively “large” microalgae (i.e., *P. micans*)
- Oligotrophic waters (i.e., Florida Keys)

In field trials, compliance tools had a high probability (~99%) of detecting an exceedance when concentrations were $\geq 30 \text{ mL}^{-1}$:

- Therefore, probabilities of detecting gross exceedances (e.g., $\geq 100 \text{ mL}^{-1}$) would be very high (~100%)

Future rounds of testing may include technologies with other approaches, new instruments, or new models of these instruments

Acknowledgements

This work was funded by the:
US Coast Guard
Research and Development Center (RDC)
(Agreement HSCGFT-14-XE51D05) and the
Maritime Administration (MARAD)

We appreciate the advice and programmatic support from
Gail Roderick (RDC), Danielle Elam (RDC), and
Carolyn Junemann (MARAD)

Acknowledgements: Participants

Additional Test Participants

Dr. Katherine 'Jenny' Carney

Smithsonian Environmental Research Center (SERC)

Ms. Lisa Allinger and Dr. Euan Reavie

University of Minnesota-Duluth; Great Ships Initiative (GSI)

Technical Advisory Committee (Fluorometry)

Dr. Ryan Albert

U.S. Environmental Protection
Agency

Dr. Richard Everett

U.S. Coast Guard

Dr. Carolyn Junemann

Maritime Administration

Dr. Sam Laney

Woods Hole Oceanographic Inst.

Dr. Beth Stauffer

University of Louisiana Lafayette



Supplemental Slides

Testing and Analysis: 2015-2016

Round 1: June – September, 2015

Round 2: March – July, 2016




Sampling ambient organisms
from seawater in Key West, FL

Accuracy: Logistical Regression

Logistical Regression compares the relationship between:

- A continuous independent variable (cell concentration)
- A binary dependent variable (Pass/Fail)

Probability
(of measuring
an exceedance)



Results: Linearity (Laboratory trials)

R^2 Values: Coefficient of Determination
Microscope counts vs. compliance tool concentrations

Tool	Laboratory Trials		
	<i>T. marina</i>	<i>P. micans</i>	Both organisms
BallastCheck2 (2015)	0.46	0.98	0.90
10Cells	0.85	0.84	0.68
YSI Ballast Monitor	0.87	0.94	0.91
BallastCheck2 (2016)	0.33	0.90	0.82
FastBallast	N/A	N/A	N/A
BW680	0.57	0.92	0.86

Legend:
 $R^2 \geq 0.90$
 $R^2 \geq 0.75$
 $R^2 < 0.50$

R^2 Values:
0 (no linear relationship) to 1 (strong linear relationship)

Results: Linearity (Field trials)

R^2 Values: Coefficient of Determination
Microscope counts vs. compliance tool concentrations

Tool	Field Trial Locations			
	NRL	GSI	SERC	All Sites
BallastCheck2 (2015)	0.63	0.64	0.12	0.10
10Cells	0.61	0.69	0.68	0.48
YSI Ballast Monitor	0.72	0.66	0.01	0.15
BallastCheck2 (2016)	0.73	0.46	0.39	0.36
FastBallast	0.13	0.75	0.71	0.37
BW680	0.66	0.61	0.82	0.66

Legend:
 $R^2 \geq 0.90$
 $R^2 \geq 0.75$
 $R^2 < 0.50$

R^2 Values:
 0 (no linear relationship) to 1 (strong linear relationship)